



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

| APPLICATION NO. | FILING DATE | FIRST NAMED INVENTOR | ATTORNEY DOCKET NO.      | CONFIRMATION NO. |
|-----------------|-------------|----------------------|--------------------------|------------------|
| 10/685,237      | 10/14/2003  | Kenneth B. Stokes    | P-3586.01 Continuation 3 | 4635             |

7590  
Kenneth J. Collier  
Medtronic, Inc.  
710 Medtronic Parkway N.E.  
Minneapolis, MN 55432

04/16/2007

|          |
|----------|
| EXAMINER |
|----------|

KAHELIN, MICHAEL WILLIAM

|          |              |
|----------|--------------|
| ART UNIT | PAPER NUMBER |
|----------|--------------|

3762

| SHORTENED STATUTORY PERIOD OF RESPONSE | MAIL DATE  | DELIVERY MODE |
|--|------------|---------------|
| 3 MONTHS                               | 04/16/2007 | PAPER         |

**Please find below and/or attached an Office communication concerning this application or proceeding.**

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

# Office Action Summary

Application No.

10/685,237

Applicant(s)

STOKES ET AL.

Examiner

Michael Kahelin

Art Unit

3762

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

## Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

## Status

- 1) ☒ Responsive to communication(s) filed on 15 November 2006.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

## Disposition of Claims

- 4) ☒ Claim(s) 36-46 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 36-46 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

## Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

## Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
  - ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

## Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/SB/08)  
Paper No(s)/Mail Date \_\_\_\_\_
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date. \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: \_\_\_\_\_

## DETAILED ACTION

### *Claim Rejections - 35 USC § 112*

1. The following is a quotation of the second paragraph of 35 U.S.C. 112:  
  
The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.
2. Claims 36-46 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.
3. In regards to claim 36, "a predetermined genetic material" is inferentially included, rendering the claim vague because it is unclear whether this element is a component of the claimed system. It is suggested to positively recite the element before further limiting it. The examiner has interpreted the claim as containing the material and should be amended accordingly.
4. In regards to claims 37 and 40, the apparatus claims recite method steps, rendering it unclear whether the claims are reciting a system or method.
5. In regards to claims 37-46, the claims are dependent on cancelled claims. The Examiner has considered all claims to depend from claim 36.
6. In regards to claims 39 and 46, the helical screw element and pacing electrode are lacking a connection to the other claimed elements. A system claim may not be a listing of parts.
7. In regards to claim 44, it is unclear what element is performing the function of "improve[ing] the ability to sense".

***Claim Rejections - 35 USC § 102***

8. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

9. Claims 36-40 and 44-46 are rejected under 35 U.S.C. 102(e) as being anticipated by Darvish et al. (US 7,190,997, hereinafter "Darvish").

10. In regards to claim 36, Darvish discloses an implantable system for delivering an ion channel protein to the heart (abstract and col. 11, line 5).

11. In regards to claims 37 and 38, the delivery means is implanted via an endocardial catheter (col. 12, line 42).

12. In regards to claim 39, the delivery system is helical (col. 6, line 29) and hollow (col. 16, line 64).

13. In regards to claim 40, the material is provided as a bolus (col. 16, line 34).

14. In regards to claims 44 and 45, the number of calcium or sodium channels is increased (col. 11, line 5), which inherently improves the ability to sense cardiac activity and increases the amplitude of the cardiac signal because more channels lead to an increased ionic current.

15. In regards to claim 46, the system comprises a sensing electrode (col. 5, line 23).

***Claim Rejections - 35 USC § 103***

16. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

17. This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

18. Claims 41-43 are rejected under 35 U.S.C. 103(a) as being unpatentable over Darvish in view of Applicant's Admitted Prior Art (hereinafter "AAPA"). Darvish discloses the essential features of the claimed invention, including genetic treatment to increase the number of sodium channel proteins (col. 11, line 5), but does not expressly disclose that this treatment is a recombinant nucleic acid molecule that encodes hH1. AAPA teaches that it is well known in the art to utilize cloned hH1 to increase sodium channel expression (page 3, line 25), thusly increasing ionic current. Therefore, it would have been obvious to one having ordinary skill in the art at the time the invention was

Art Unit: 3762

made to utilize cloned hH1 to increase sodium channel expression, thusly increasing ionic current.

### ***Double Patenting***

19. The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. A nonstatutory obviousness-type double patenting rejection is appropriate where the conflicting claims are not identical, but at least one examined application claim is not patentably distinct from the reference claim(s) because the examined application claim is either anticipated by, or would have been obvious over, the reference claim(s). See, e.g., *In re Berg*, 140 F.3d 1428, 46 USPQ2d 1226 (Fed. Cir. 1998); *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) or 1.321(d) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent either is shown to be commonly owned with this application, or claims an invention made as a result of activities undertaken within the scope of a joint research agreement.

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

20. Claims 36-46 are rejected on the ground of nonstatutory obviousness-type

double patenting as being unpatentable over claims 1-11 of U.S. Patent No. 6,567,705.

Although the conflicting claims are not identical, they are not patentably distinct from each other because patent's claims are more narrow and meet the limitations of the broader claims of this application.

***Conclusion***

21. Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael Kahelin whose telephone number is (571) 272-8688. The examiner can normally be reached on M-F, 9-5.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Angela Sykes can be reached on (571) 272-4955. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3762

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

MWK

*MWK*  
4/6/07

*GE*  
GEORGE R. EVANICKO  
PRIMARY EXAMINER

4/12/07